

Title: Clinical evaluation of the Diagnostic Analyzer for Selective Hybridization (DASH): a point-of-care PCR test for rapid detection of SARS-CoV-2 infection

Running Title: Clinical evaluation of DASH for SARS-CoV-2 infection

Authors

Chad J Achenbach^{1,2,3}, Matthew Caputo¹, Claudia Hawkins^{1,2}, Lauren C Balmert³, Chao Qi⁴, Joseph Odorisio¹, Etienne Dembele¹, Alema Jackson⁵, Hiba Abbas⁵, Jennifer K Frediani^{6,7}, Joshua M Levy^{6,8}, Paulina A. Rebolledo^{6,9}, Russell R. Kempker^{6,9}, Annette M. Esper^{6,10}, Wilbur A. Lam, MD, PhD^{6,11}, Greg S. Martin^{6,10}, and Robert L Murphy^{1,2}

Affiliations

¹Havey Institute for Global Health, Feinberg School of Medicine, Northwestern University;

²Department of Medicine, Division of Infectious Diseases, Feinberg School of Medicine,

Northwestern University; ³Department of Preventive Medicine, Feinberg School of Medicine,

Northwestern University; ⁴Department of Pathology, Northwestern University; ⁵Access

Community Health Network; ⁶Atlanta Center for Microsystems Engineered Point-of-Care

Technologies; ⁷Emory University Nell Hodgson Woodruff School of Nursing; ⁸Emory University

Department of Otolaryngology; ⁹Emory University Division of Infectious Diseases; ¹⁰Emory

University Division of Pulmonary, Allergy, Critical Care and Sleep Medicine; ¹¹Emory University

Department of Pediatrics

Corresponding Author: Chad J. Achenbach, M.D., M.P.H., Havey Institute for Global Health, Feinberg School of Medicine, Northwestern University, 710 N Lake Shore Drive, Chicago, IL 60611, email: c-achenbach@northwestern.edu, phone: 312-503-8810, fax: 312-503-8800

Alternate Corresponding Author: Robert L. Murphy, M.D., Havey Institute for Global Health, Feinberg School of Medicine, Northwestern University, 710 N Lake Shore Drive, Chicago, IL 60611, email: r-murphy@northwestern.edu, phone: 312-503-9000, fax: 312-503-8800

Word Count Text: 2,929

Word Count Abstract: 327

Key Words: SARS-CoV-2, COVID-19, PCR, testing

Summary: DASH is an accurate, easy to use, and fast point-of-care test with applications for diagnosis and screening of SARS-CoV-2 infection.

Abstract

Background: Rapid and accurate testing for SARS-CoV-2 is an essential tool in the medical and public health response to the COVID-19 pandemic. An ideal test for COVID-19 would combine the sensitivity of laboratory-based PCR combined with the speed and ease of use of point-of-care (POC) or home-based rapid antigen testing.

Methods: To evaluate the performance of the Diagnostic Analyzer for Selective Hybridization (DASH) SARS-CoV-2 POC PCR (sample insertion to result time of 16 minutes), we conducted a cross-sectional study of adults with and without symptoms of COVID-19 at four clinical sites. We collected two bilateral anterior nasal swabs from each participant and information on COVID-19 symptoms, vaccination, and exposure. One swab was tested with the DASH SARS-CoV-2 POC PCR and the second in a central laboratory using Cepheid Xpert Xpress SARS-CoV-2 PCR. We assessed test concordance and calculated sensitivity, specificity, negative and positive predictive values using Xpert as the “gold standard.”

Results: We enrolled 315 and analyzed 313 participants with median age 42 years; 65% were female, 62% symptomatic, 75% had received ≥ 2 doses of mRNA COVID-19 vaccine, and 16% currently COVID-19 positive. There were concordant results for 307 tests indicating an overall agreement for DASH of 0.98 [95% CI 0.96, 0.99] compared to Xpert. DASH performed at 0.96 [95% CI 0.86, 1.00] sensitivity and 0.98 [95% CI 0.96, 1.00] specificity, with a positive predictive value of 0.85 [95% CI 0.73, 0.96] and negative predictive value of 0.996 [95% CI 0.99, 1.00]. The six discordant tests between DASH and Xpert all had high Ct values (>30) on the respective positive assay. DASH and Xpert Ct values were highly correlated ($R=0.89$ [95% CI 0.81, 0.94]).

Conclusions: DASH POC SARS-CoV-2 PCR was accurate, easy to use, and provided fast results in real-life clinical settings with an overall performance similar to an EUA-approved laboratory-based PCR. Its compact design and ease of use are optimal for a variety of healthcare, and potentially community settings, including areas with lack of access to central laboratory-based PCR testing.

Introduction

Rapid and accurate detection of SARS-CoV-2 infection has been a key component of the medical and public health response to the COVID-19 pandemic[1–3]. Despite advances in testing, we need better tools for early diagnosis and screening, particularly as we move toward “test and treat” strategies given several outpatient therapies for COVID-19 have been granted FDA emergency use authorization[4–7]. To date, the “gold standard” for accuracy, with highest analytic sensitivity, are nucleic acid amplification test (NAAT)-based technologies such as RT-PCR[8,9]. However, PCR testing requires centralized laboratory equipment, a high level of technical expertise, is expensive, time consuming, and is unavailable in many rural or under-resourced settings. Current rapid point-of-care (POC) and home tests are antigen or NAAT non-PCR-based technologies are simple to use, but with lower sensitivity resulting in a limited ability to detect individuals early in SARS-CoV-2 infection when pre-symptomatic or asymptomatic[10–23]. An ideal test for COVID-19 would have the accuracy of laboratory-based PCR testing combined with the speed and ease of use of POC rapid antigen testing[1].

To address this need, Minute Molecular Diagnostics, Inc. (m2dx.com), with support from the U.S. National Institutes of Health (NIH) National Institute of Biomedical Imaging and Bioengineering (NIBIB) Rapid Acceleration of Diagnostics (RADxSM Tech) program (<https://www.nibib.nih.gov/covid-19/radx-tech-program>), developed the Diagnostic Analyzer for Selective Hybridization (DASH), a sample-to-answer platform designed for POC COVID-19 testing in healthcare settings with potential applications in the community. DASH utilizes rapid PCR and microfluidic fabrication technology with uniquely designed cartridges and an analyzer that are easy to use, do not require technical expertise or specialized training, and produce results in approximately 15 minutes. In laboratory testing, DASH was found to have analytic sensitivity at the same level as laboratory-based PCR with the ability to detect at least 100 copies/mL of SARS-CoV-2 RNA.

The primary objective of this clinical validation study was to test the performance and accuracy of the DASH POC test for diagnosis of SARS-CoV-2 infection among individuals with and without COVID-19 symptoms in outpatient settings.

Methods

Study Design, Population and Recruitment

We performed a cross-sectional study from June to September 2021 of adult (18 years and older) participants with or without symptoms of COVID-19 who were tested for SARS-CoV-2 at one of four outpatient clinical sites. Asymptomatic participants self-reported no symptoms attributable to COVID-19 within 14 days prior to enrollment; however, prior to August 10, 2021, several participants were enrolled under previous FDA asymptomatic criteria of no symptoms within 48 hours. Symptomatic participants self-reported symptoms attributable to COVID-19 within the 24 hours prior to enrollment that started within the prior seven days. Symptoms attributable to COVID-19 include the following (as designated by CDC): fever or chills, cough, shortness of breath or difficulty breathing, fatigue, muscle or body aches, headache, new loss of taste or smell, sore throat, congestion or runny nose, nausea or vomiting, and diarrhea[24]. Participants were consented remotely via telephone utilizing Northwestern REDCap and emailed a survey link for data collection on symptoms, exposure details, demographics, vaccine history, risk factors for acquisition, and co-morbidities.

We had three phases of recruitment. Initially, from June to July 2021, we prospectively recruited and enrolled participants only at Chicago sites (Northwestern Memorial Hospital (NMH) Infectious Diseases Center (IDC) and NMH Clinical Research Unit (CRU)) regardless of risk or exposure and without recent known testing. Second, due to low COVID-19 prevalence in Chicago, in August 2021, we aimed to enhance test positivity by focusing recruitment on younger unvaccinated participants (18 to 45 years) and those with high-risk COVID-19 exposures, again only at Chicago sites (NMH IDC, NMH CRU, and Access Community Health Network). Finally, in September 2021, we performed enhanced recruitment where we included

participants from a high COVID-19 prevalence outpatient sites in Georgia who were locally enrolled and consented in the Emory University “RADxtra COVID-19 Test Verification” study. Under our enhanced recruitment, we allowed for enrollment of individuals within seven days of testing positive for SARS-CoV-2 by a PCR-based assay. Clinical information from participants enrolled at the Emory RADxtra site was limited to data obtained through their study protocol.

Sample Collection and Testing

After we obtained informed consent and participants completed REDCap-based on-line surveys, we performed a single research visit at each clinical site. At the visit, we obtained two bilateral anterior nasal swabs from each participant. To assure each swab was obtained from a “fresh” nostril, the study team clinician (physician or research nurse) first swabbed the left nostril anterior nares by inserting the swab at least 1 cm (0.5 inch) swirling the swab around the nasal wall for 10-15 seconds, then switched to the second swab and sampled the right nostril (same procedure as above), used the second swab to sample the left nostril, and finally switched back to the first swab to sample the right nostril. The first swab was placed into a designated plastic sleeve for testing on DASH SARS-CoV-2 PCR (DASH) machine (Figure 1A) and the second swab placed into viral transport media (VTM), lightly rotated for a few seconds, broken off, and sent for Cepheid Xpert Xpress SARS-Cov-2 RT-PCR (Xpert) testing. The DASH swab was inserted into the DASH cartridge and run on the machine as per the Quick Reference Guide (QRG) provided by Minute Molecular Diagnostics, Inc. at POC within 30 minutes of collection (Figure 1B). The nasal swab in VTM was transported to the Northwestern Memorial Hospital (NMH) clinical microbiology laboratory and run on the Xpert testing platform within 4 hours of collection. Participants were given the results of only the Xpert standard PCR testing within 24 hours. If DASH and Xpert testing were discordant, then if residual VTM from Xpert testing was available, we performed tie breaker testing of residual VTM using Abbott Alinity M platform.

For research visits from May 22 to August 6, 2021, DASH testing machines were not available at the POC. During this period, the DASH nasal swab was placed into -80°C freezer

in the NMH clinical microbiology laboratory within 4 hours of collection for temporary storage. Once DASH machines became available at POC, then frozen DASH samples were thawed to room temperature for at least an hour (no more than 4 hours) and placed into the DASH cartridge and assay run as per the QRG. Prior laboratory experiments found nasal swabs to be stable under these conditions with less than 6% difference in virus yield (see Supplement).

The RADxtra study team at Emory also collected two bilateral anterior nasal swabs as per the same study procedure detailed above and then shipped de-identified DASH swabs and corresponding VTM overnight to Chicago on ice packs (2-8°C) in compliance with the Cepheid Xpert Xpress SARS-CoV-2 Emergency Use Authorized IFU[25]. Upon delivery to Northwestern, the direct nasal swab was run on the DASH machine as per QRG and the VTM nasal swab specimen was run on Xpert per the manufacturer instructions in the NM clinical microbiology laboratory. Again, prior laboratory experiments found nasal swabs to be stable under these conditions with less than 6% difference in virus yield (see Supplement). We provided RADxtra investigators at Emory a unique identifier for each participant of samples/data sent to Northwestern. They kept this link to their study ID locally at Emory and none of the investigators or team at Northwestern had access to this linkage.

A usability survey was distributed to the DASH users after each use of the machine. This survey included questions on DASH complexity and ease of use with no specific training and only instruction from the QRG. For this initial study, we only analyzed the survey response from users who completed the survey after their first use of the DASH machine.

Statistical Analysis

Descriptive statistics were used to summarize demographics, exposure, and COVID-19 risk factors. Primary analyses calculated sensitivity and specificity from a 2X2 contingency table comparing DASH to the Xpert PCR. Specifically, sensitivity was calculated as the $\Pr(\text{DASH+}|\text{Xpert+})$ or True Positives / (True Positives + False Negatives). Specificity was calculated as the $\Pr(\text{DASH-}|\text{Xpert-})$ or True Negatives / (True Negatives + False Positives). In

secondary analyses, we also calculated positive predictive value (PPV) and negative predictive value (NPV) using Bayes theorem assuming an *a priori* specified prevalence of 8%. Given the changing dynamic of the pandemic over the course of the study, we also estimated PPV and NPV for a range of prevalence estimates (1% to 30%). Confidence intervals for predictive values were generated using a percentile bootstrap, with 1000 bootstrap replicates. Additionally, positive likelihood ratio (LR+) and negative likelihood ratio (LR-) were calculated as sensitivity/(1-specificity) and (1-sensitivity)/specificity, respectfully. 95% confidence intervals (CIs) were calculated for all estimates. Descriptive statistics were also used to summarize the DASH usability survey.

Within the subset of participants that tested positive on both Xpert and DASH, a Pearson correlation coefficient was used to assess the association between DASH and Xpert Ct values. A Bland-Altman plot visualized agreement between the two measures, using percentiles for the 95% CI to account for non-normality of differences.

All analyses were performed using R 4.1.1 (R Core Team, 2021) and the epiR (v2.0.36; Sergeant & Stevenson, 2021) package.

Sample Size Calculations

Sample size considerations were based on estimates of sensitivity and specificity, with corresponding 95% CIs, necessary to meet minimum FDA performance criteria for Emergency Use Authorization (EUA). Specifically, a sample size of at least 275 participants was required to estimate sensitivity of 0.95 such that the lower bound of the 95% CI was greater than 0.76, and to estimate specificity of 0.98 such that the lower bound of the CI was greater than 0.95.

Calculations assumed a population prevalence of positive SARS-CoV-2 of 8% and at least 30 participants overall in the study with a positive Xpert “gold standard” test result.

Results

We enrolled 315 participants and obtained valid results from 313 participants at our four clinical sites with samples run by 15 different users on three DASH machines. Among our 313 participants analyzed, median age was 42 years, 65% were female, 75% had received two doses of mRNA COVID-19 vaccine (Pfizer or Moderna), 62% had COVID-19 symptoms, 14% had known exposure to someone COVID-19, and 16% had known SARS-CoV-2 PCR positive testing prior to study enrollment (Table 1). Excluding participants from the Emory RADextra site with limited clinical data (n=116), 27% had co-morbidities.

We found a total of 51 participants tested positive on DASH and 49 participants tested positive on Xpert (Table 2). Summaries of DASH compared to Xpert testing are reported in Tables 2 and 3. Sensitivity of the DASH test was estimated at 0.96 [95% CI: 0.86, 1.00] and specificity was estimated at 0.98 [95% CI: 0.96, 1.00]. There was concordance of 307 tests indicating an overall agreement of 0.98 [95% CI: 0.96, 0.99]. The DASH test had a PPV of 0.85 [95% CI: 0.73, 0.96] and NPV of 0.996 [95% CI: 0.99, 1.00] as compared to the Xpert. Predictive value estimates for a range of infection prevalence rates are depicted in Figure 2 to assess performance across a wide range of COVID-19 prevalence. At the observed disease prevalence within the sample (16%), DASH demonstrated a PPV of .92 [.86, .98] and NPV of .99 [.98, 1.00], however this prevalence is likely an overestimate as known-positive participants were enrolled under our enhanced recruitment strategy. Within the subgroup of participants who tested positive on both Xpert and DASH (n=45), Ct values for DASH and Xpert were highly correlated (Pearson correlation coefficient = 0.89 [95% CI .81, .94], Figure 3A). Bland-Altman plot indicated a negative bias, such that Xpert Ct values were higher on average by 7.9 cycles (Figure 3B).

Clinical and testing information for individual participants with discordant results (n=6) are summarized in Table 4. We had two participants with negative DASH testing and positive Xpert testing and four participants with positive DASH testing and negative Xpert testing. All six of these individuals had at least one COVID-19 symptom and/or known positive PCR testing for COVID prior to enrollment. Both false negatives had high Ct values (>35 cycles) on Xpert testing, and 3 of the 4 false positives had high Ct values (>35 cycles) or an indeterminate result (one of two SARS-CoV-2 gene targets detected) on the tiebreaker assay. Five of these individuals with discordant results were late in their course of infection (at least 4 days from onset of symptoms or prior positive PCR testing).

Ten of fifteen users completed our usability survey after their first time running a participant sample on the DASH machine. When asked to rate their agreement with the statement “I found this DASH test was easy to use”, 40% strongly agreed (4 users), 50% agreed (5 users), and one user was neutral. When asked, “would you recommend DASH test to someone else?”, 50% responded “definitely yes” (5 users) 40% responded “probably yes” (4 users), and one user did not respond.

Discussion

In this cross-sectional clinical study, we found that the DASH SARS-CoV-2 PCR was easy to use, fast (time from placement of direct nasal swab into cartridge until result readout of 16 minutes) and performed with excellent test characteristics (sensitivity of 0.96, specificity of 0.98, and overall agreement of 0.98) compared to a “gold standard” laboratory-based PCR in real-life outpatient healthcare settings. From 313 participants with valid results, there were only six discordant results and each of these participants had at least one clinical symptom of COVID-19 or known positive PCR tests. DASH was negative and missed only two infected individuals; however, they both had high Ct values (>39) on standard PCR and thus were unlikely to be contagious[26–29]. In contrast, DASH detected four individuals with SARS-CoV-2 that were missed by Xpert, with only one of these as a potential false positive as RNA was not

detected on tie breaker PCR testing and the participant only had dyspnea without other symptoms typical of COVID-19. Discordance only occurred at high Ct values and is consistent with known limitations of SARS-CoV-2 PCR testing assays early or late in infection when virus levels are low[8,9,30,31]. Thus, we believe that DASH performed equivalent to laboratory-based PCR testing and displays its potential as a rapid and highly sensitive POC option to supplement or replace current PCR platforms in real-life clinical settings and eventually the community.

DASH improves upon current POC non-PCR NAAT technologies for SARS-CoV-2 such as Abbott ID NOW. In several studies, the sensitivity of the ID NOW test was found to be between 0.75 – 0.98 depending on number of low positive samples (Ct value >35 on standard PCR assays)[32–37]. Limiting to positive Xpert with Ct values from 35 to 39, DASH detected all low positive samples, thus displaying an ability to potentially detect virus earlier in the course of SARS-CoV-2 infection compared to ID NOW. This has become an important public health issue when testing highly vaccinated populations and those with emerging variants, such as Omicron, who may have a shorter window of SARS-CoV-2 detection or lower levels of virus after high risk exposure[38–41].

In addition, DASH is easy to use (Clinical Laboratory Improvement Amendments (CLIA)-waived) and does not require technical or laboratory training to run the machine. Our research team of 15 different users had neither formal training nor laboratory expertise and were able to perform DASH testing with valid results from 99.4% of participants sampled. Only using instruction from the QRG, 90% of our users assessed DASH to be easy to use and they all recommended DASH to others who might be doing COVID testing. We plan to expand these assessments as more users have access to DASH and perform qualitative research to optimize DASH usability in clinical and non-clinical settings.

Correlation analyses (see Figure 2A) between DASH and Xpert Ct values demonstrated a strong association; however, Bland-Altman plots (Figure 2B) indicated DASH PCR Ct values

were systematically lower than Xpert PCR. These differences could occur for several possible technical reasons. DASH and Xpert PCRs both detect two SARS-CoV-2 targets; however, DASH does this in the same fluorescent channel whereby Xpert splits molecular material between two separate channels. Also, for DASH, the nasal swab was directly inserted into the cartridge whereas Xpert swab was placed in 3 mL of VTM, and then 300 µl of liquid media added to the cartridge. Finally, DASH processes approximately 50% of the specimen and Xpert only processes 10%. These results suggest that DASH can provide meaningful semi-quantitative viral load measurements, similar to laboratory-based PCR, but at POC and quick enough to potentially influence clinical or public health guidance.

Our study had several limitations. First, the cross-sectional design did not allow us to understand test characteristics of DASH over time and throughout the dynamic range of SARS-CoV-2 infection within individuals. Second, most of our participants were symptomatic and only one positive sample was from an asymptomatic participant. Additionally, our participants were young, highly vaccinated, and without major co-morbidities including immune suppression. Thus, DASH test characteristics are less certain for those at greatest risk for severe COVID-19 or among certain special populations. Although, other PCR-based SARS-CoV-2 testing technologies have generally performed well across many different high-risk patient groups[42–47].

DASH is an important advancement in molecular POC diagnostics for SARS-CoV-2 with potential applications beyond many non-PCR NAAT technologies. DASH could decentralize PCR diagnostics by bringing molecular testing to a wider range of community settings in the U.S. and throughout the world. We need highly sensitive and easy to use technologies with the ability to quantify virus for monitoring SARS-CoV-2 infection dynamics and when to safely end isolation. Further research will allow us to understand how DASH could be utilized for detecting emerging SARS-CoV-2 variants, community surveillance, or large non-health care screening of travelers, students, educators, and front-line workers.

Funding

This work was supported by the National Institutes for Health (NIH)/National Institute for Biomedical Imaging and Bioengineering (NIBIB)(U54EB027049-02S1 and U54EB027690). We utilized Northwestern REDCap which is supported by Northwestern University Clinical and Translational Sciences Institute (NUCATS). NUCATS is funded in part by a Clinical and Translational Science Award (CTSA) grant from the NIH (UL1TR001422). The Emory RADxtra clinics are supported by NIH NIBIB (U54EB027690) Emory University's Children's Clinical and Translational Discovery Core and Children's Healthcare of Atlanta, Grady Health System.

Authorship Contributions and Conflict of Interest

C.J.A., R.L.M., C.H., L.C.B., and C.Q. designed the study, wrote the research study protocol, standard operating procedures, and *a priori* statistical analysis plans. M.C., J.O., and E.D. conducted participant recruitment, enrollment, consenting, and scheduling at NMH IDC and CRU sites. C.J.A. and C.H. performed study visits and obtained nasal swabs samples at NMH IDC. A.J. and H.A. conducted participant recruitment, enrollment, consenting, and study visits including nasal swabs at Access Community Health Network. J.K.F., J.M.L., P.A.R., R.R.K., A.M.E., W.A.L., and G.S.M. conducted participant recruitment, enrollment, consenting, and study visits including nasal swabs at Emory RADxtra clinics. Statistical analyses were performed by M.C. with assistance from L.C.B. and C.J.A. The initial draft of the manuscript was written by C.J.A. and M.C. Throughout the study, all authors participated in discussions about the study conduct, analyses, and interpretation of findings. All authors were also involved in the review and editing process of the final manuscript for submission. Minute Molecular Diagnostics, Inc. provided DASH machines, cartridges, Puritan nasal swabs, and technical support (including providing Supplement); however, they were not involved in the research design, participant recruitment, enrollment, informed consent, study visits, sample collection, DASH/Xpert testing, data collection, statistical analyses, drafting or editing of this manuscript. All authors declare that they have no conflicts of interest.

Acknowledgement

These findings are presented on behalf of the study participants who generously gave their time, samples, and information for this research. We would also like to thankfully acknowledge all staff and faculty in the NMH IDC, NMH CRU, Access Community Health Network, Emory RADxtra clinics, and NMH microbiology laboratory who contributed to this study. Emory investigators would like to acknowledge the following individual team members for their assistance with this study: Tamara Wesley, Tim Thurman, Kristi Godbolt, Anna Wood, Adrianna Westbrook, Leona Wells, Julie Sullivan, Cheryl L. Stone, and Jared O'Neal. Northwestern and Access investigators would like to acknowledge the following individual team members for their assistance with this study: Kate Klein, Camille Bundy, Mackenzie Furnari-Stickney, Amelia Kelly, Jehannaz Dastoor, Michael Govern, Carol Govern, Kristen Weber, Elizabeth Christian, Adrianna Quintana, Patricia Helbin, and Vinay Durbhakula.

Table 1. Characteristics of 313 participants with a valid Diagnostic Analyzer for Selective Hybridization (DASH) SARS-CoV-2 PCR test result.

Characteristic	N (%) or Median (IQR)
Clinical Site	
Total	313 (100)
NMH IDC	112 (36)
NMH CRU	71 (23)
Access Community Health Network	14 (5)
Emory RADextra Clinics	116 (37)
Age in years	42 (32, 57)
Sex Assigned at Birth	
Female	204 (65)
Male	107 (34)
Not Reported	2 (1)
Race*	
White	155 (79)
Asian	17 (9)
Black or African American	17 (9)
Other or unknown	8 (4)
Ethnicity*	
Non-Hispanic or unknown	187 (95)
Hispanic	10 (5)
COVID-19 Vaccine Doses	
0	55 (18)
1	23 (7)
2	229 (73)
Vaccinated unknown doses	6 (2)
Vaccine Manufacturer****	
Pfizer	145 (56)
Moderna	89 (34)
Johnson & Johnson	15 (6)
Unknown type	9 (4)
Days since last vaccine dose****	140 (101, 171)
Samples from Enrichment	51 (16)
Known COVID-19 Exposure within 2 weeks	28 (9)

Known COVID-19 Positive (PCR positive within 7 days)	50 (16)
COVID-19 Symptomatic	
Yes	195 (62)
No	118 (38)
Individual COVID-19 Symptoms**	
Congestion	78 (40)
Cough	78 (40)
Headache	68 (35)
Fatigue	60 (31)
Fever/Chills	59 (30)
Sore/Scratchy Throat	53 (27)
Vomiting/Nausea/Diarrhea	51 (26)
Myalgias	49 (25)
Loss of Taste/Smell	37 (19)
Shortness of Breath	25 (13)
Arthralgias	25 (13)
Abdominal Pain	20 (10)
Photophobia	15 (8)
Medical History*	
Any co-morbidity	52 (26)
Hypertension	22 (11)
Asthma	9 (5)
Diabetes	7 (4)
Cancer	6 (3)
Immunodeficiency (Not HIV)	5 (3)
Coronary Artery Disease	4 (2)
Anemia	4 (2)
Other***	9 (5)

*Excluding 116 participants from Emory RADxtra site as they did not provide this demographic or clinical information (n=197)

**Among those reporting symptoms (n=195); categories are not mutually exclusive

***Others were COPD (n=3), renal failure (n=2), stroke/TIA (n=2), HIV (n=1), and MI (n=1)

****Among those that received at least 1 vaccine dose prior to testing (n=258)

Table 2. Summary of Diagnostic Analyzer for Selective Hybridization (DASH) PCR results compared to Cepheid Xpert Xpress (Xpert) PCR results.

	Xpert Positive	Xpert Negative	Total
DASH Positive	47	4*	51
DASH Negative	2*	260	262
Total	49	264	313

*See Table 4 for summary of clinical and testing details from 6 participants with discordant DASH and Xpert results

Table 3. Diagnostic Analyzer for Selective Hybridization (DASH) POC PCR diagnostic performance and accuracy compared to Cepheid Xpert Xpress PCR as “gold standard.”

<i>Statistic</i>	<i>n/N</i>	<i>Estimate [95% CI]</i>
Apparent Prevalence	49/313	0.16 [0.12, 0.21]
Sensitivity	47/49	0.96 [0.86, 1.00]
Specificity	260/264	0.98 [0.96, 1.00]
Overall Diagnostic Accuracy	307/313	0.98 [0.96, 0.99]
Positive Predictive Value*		0.85 [0.73, 0.96]
Negative Predictive Value*		0.996 [0.99, 1.00]
Positive Likelihood Ratio		63.31 [23.90, 167.71]
Negative Likelihood Ratio		0.04 [0.01, .16]

*Estimated using Bayes' theorem assuming a prevalence of 8%

Table 4. Summary of Diagnostic Analyzer for Selective Hybridization (DASH) PCR and Cepheid

Xpert Xpress (Xpert) PCR discordant results.

Discordant Results and Tiebreakers

<i>Number, Site</i>	<i>Xpert PCR (Ct if positive)</i>	<i>DASH PCR (Ct if positive)</i>	<i>Tiebreaker Result (Ct if positive) *</i>	<i>Known Positive Referral Platform (days prior, Ct)</i>	<i>Symptoms; Onset; Vaccine doses (type)</i>
1, Emory	Detected (42.5)	Not Detected	Not Done; sample discarded	Roche Cobas 6800 (unknown)	Cough, Headache, Sore throat, Fatigue, Diarrhea, Shortness of breath; 4 days; Unvaccinated
2, NMH ID	Detected (39.8)	Not Detected	Not Done; sample discarded	Abbott Alinity M (5, 34.1)	No symptoms; NA; 2 doses (Pfizer)
3, NMH ID	Not Detected	Detected (34.1)	Indeterminate	Cepheid Xpert Xpress (2, 38.2)	Cough, sinus congestion, loss of taste and smell; 17 days; 1 dose (J&J)
4, NMH ID	Not Detected	Detected (29.9)	Detected (36.2)	Thermo Fisher TaqPath COVID-19 Combo Kit (8, Ct unknown)	Scratchy throat, Sore throat, Cough, Runny nose, Fever/chills, Loss of taste or smell; 13 days; 1 dose (J&J)
5, Emory	Not Detected	Detected (33.2)	Detected (36.8)	Not done	Headache, Sore throat, Abdominal pain, Loss of sense of taste or smell; 4 days; 2 doses (Pfizer)
6, Emory	Not Detected	Detected (33.4)	Not Detected	Abbott Alinity M (unknown)	Shortness of breath; 1 day; 2 doses (Pfizer)

**All tiebreaker assays run on Abbott Alinity M platform*

Figure 1. Diagnostic Analyzer for Selective Hybridization (DASH) PCR machine at point-of-care in clinical space (panel A) and workflow schematic (panel B): barcode on cartridge scanned, nasal swab added to cartridge and snapped at break point, cartridge loaded into DASH machine, and reported generated on screen/printed after approximately 15 minutes.

Figure 1A



Figure 1B




Figure 2. Positive (red line) and negative (green line) predictive values of Diagnostic Analyzer for Selective Hybridization (DASH) SARS-CoV-2 PCR test by varying COVID-19 prevalence.




Figure 3. Correlation (panel A) and Bland-Altman (panel B) analyses for comparison of Diagnostic Analyzer for Selective Hybridization (DASH) PCR and Cepheid Xpert Xpress PCR cycle threshold (Ct) values.

Figure 3A





Figure 3B



References

1. Drain PK. Rapid Diagnostic Testing for SARS-CoV-2. *New Engl J Med*. 2022. doi:10.1056/nejmcp2117115
2. Diagnostic testing for SARS-CoV-2. [cited 24 Jan 2022]. Available: <https://www.who.int/publications/i/item/diagnostic-testing-for-sars-cov-2>
3. Overview of Testing for SARS-CoV-2, the virus that causes COVID-19 | CDC. [cited 24 Jan 2022]. Available: <https://www.cdc.gov/coronavirus/2019-ncov/hcp/testing-overview.html>
4. Mahase E. Covid-19: Pfizer's paxlovid is 89% effective in patients at risk of serious illness, company reports. *Bmj*. 2021;375: n2713. doi:10.1136/bmj.n2713
5. Ledford H. COVID antiviral pills: what scientists still want to know. *Nature*. 2021;599: 358–359. doi:10.1038/d41586-021-03074-5
6. Bernal AJ, Silva MMG da, Musungaie DB, Kovalchuk E, Gonzalez A, Reyes VD, et al. Molnupiravir for Oral Treatment of Covid-19 in Nonhospitalized Patients. *New Engl J Med*. 2021; NEJMoa2116044. doi:10.1056/nejmoa2116044
7. Weinreich DM, Sivapalasingam S, Norton T, Ali S, Gao H, Bhore R, et al. REGEN-COV Antibody Combination and Outcomes in Outpatients with Covid-19. *New Engl J Medicine*. 2021;385: NEJMoa2108163. doi:10.1056/nejmoa2108163
8. Ash C. Added value of PCR testing for COVID-19. *Science*. 2021;373: 291.12-293. doi:10.1126/science.373.6552.291-l
9. Garg A, Ghoshal U, Patel SS, Singh DV, Arya AK, Vasanth S, et al. Evaluation of seven commercial RT-PCR kits for COVID-19 testing in pooled clinical specimens. *J Med Virol*. 2021;93: 2281–2286. doi:10.1002/jmv.26691
10. Allan-Blitz L-T, Klausner JD. A Real-World Comparison of SARS-CoV-2 Rapid Antigen Testing versus PCR Testing in Florida. *J Clin Microbiol*. 2021;59: e01107-21. doi:10.1128/jcm.01107-21
11. Kendall EA, Arinaminpathy N, Sacks JA, Manabe YC, Dittrich S, Schumacher SG, et al. Antigen-based Rapid Diagnostic Testing or Alternatives for Diagnosis of Symptomatic COVID-19. *Epidemiology*. 2021;32: 811–819. doi:10.1097/ede.0000000000001400
12. Yamayoshi S, Sakai-Tagawa Y, Koga M, Akasaka O, Nakachi I, Koh H, et al. Comparison of Rapid Antigen Tests for COVID-19. *Viruses*. 2020;12: 1420. doi:10.3390/v12121420
13. Dinnes J. COVID-19 rapid antigen testing strategies require careful evaluation. *Ebiomedicine*. 2021;70: 103491. doi:10.1016/j.ebiom.2021.103491

14. Loeffelholz MJ, Tang Y-W. Detection of SARS-CoV-2 at the point of care. *Bioanalysis*. 2021;13: 10.4155/bio-2021-0078. doi:10.4155/bio-2021-0078
15. Okoye GA, Kamara HI, Strobeck M, Mellman TA, Kwagyan J, Sullivan A, et al. Diagnostic accuracy of a rapid diagnostic test for the early detection of COVID-19. *J Clin Virol*. 2021; 105023. doi:10.1016/j.jcv.2021.105023
16. Jegerlehner S, Suter-Riniker F, Jent P, Bittel P, Nagler M. Diagnostic accuracy of a SARS-CoV-2 rapid antigen test in real-life clinical settings. *Int J Infect Dis*. 2021;109: 118–122. doi:10.1016/j.ijid.2021.07.010
17. Mak GC, Cheng PK, Lau SS, Wong KK, Lau C, Lam ET, et al. Evaluation of rapid antigen test for detection of SARS-CoV-2 virus. *J Clin Virol*. 2020;129: 104500–104500. doi:10.1016/j.jcv.2020.104500
18. Matsuda EM, Campos IB, Oliveira IP, Colpas DR, Carmo AM dos S, Brígido LF de M. Field evaluation of COVID-19 antigen tests versus RNA based detection: Potential lower sensitivity compensated by immediate results, technical simplicity, and low cost. *J Med Virol*. 2021;93: 4405–4410. doi:10.1002/jmv.26985
19. Scohy A, Anantharajah A, Bodéus M, Kabamba-Mukadi B, Verroken A, Rodriguez-Villalobos H. Low performance of rapid antigen detection test as frontline testing for COVID-19 diagnosis. *J Clin Virol*. 2020;129: 104455–104455. doi:10.1016/j.jcv.2020.104455
20. Mitchell SL, Orris S, Freeman T, Freeman MC, Adam M, Axe M, et al. Performance of SARS-CoV-2 antigen testing in symptomatic and asymptomatic adults: a single-center evaluation. *Bmc Infect Dis*. 2021;21: 1071. doi:10.1186/s12879-021-06716-1
21. García-Fiñana M, Buchan IE. Rapid antigen testing in COVID-19 responses. *Science*. 2021;372: 571–572. doi:10.1126/science.abi6680
22. Harmon A, Chang C, Salcedo N, Sena B, Herrera BB, Bosch I, et al. Validation of an At-Home Direct Antigen Rapid Test for COVID-19. *Jama Netw Open*. 2021;4: e2126931. doi:10.1001/jamanetworkopen.2021.26931
23. Smith RL, Gibson LL, Martinez PP, Ke R, Mirza A, Conte M, et al. Longitudinal assessment of diagnostic test performance over the course of acute SARS-CoV-2 infection. *J Infect Dis*. 2021;224: jia3337-. doi:10.1093/infdis/jia3337
24. Symptoms of COVID-19 | CDC. [cited 24 Jan 2022]. Available: <https://www.cdc.gov/coronavirus/2019-ncov/symptoms-testing/symptoms.html>
25. Cepheid. Xpert Xpress SARS-CoV-2: Instructions for Use Under and Emergency Use Authorization (EUA) Only. 2021. Available: <https://www.fda.gov/media/136314/download>
26. Meyerowitz EA, Richterman A, Gandhi RT, Sax PE. Transmission of SARS-CoV-2: A Review of Viral, Host, and Environmental Factors. *Ann Intern Med*. 2020;174: M20-5008. doi:10.7326/m20-5008

27. Singanayagam A, Hakki S, Dunning J, Madon KJ, Crone MA, Koycheva A, et al. Community transmission and viral load kinetics of the SARS-CoV-2 delta (B.1.617.2) variant in vaccinated and unvaccinated individuals in the UK: a prospective, longitudinal, cohort study. *Lancet Infect Dis.* 2021. doi:10.1016/s1473-3099(21)00648-4
28. Singanayagam A, Patel M, Charlett A, Bernal JL, Saliba V, Ellis J, et al. Duration of infectiousness and correlation with RT-PCR cycle threshold values in cases of COVID-19, England, January to May 2020. *Eurosurveillance.* 2020;25: 2001483. doi:10.2807/1560-7917.es.2020.25.32.2001483
29. Wölfel R, Corman VM, Guggemos W, Seilmaier M, Zange S, Müller MA, et al. Virological assessment of hospitalized patients with COVID-2019. *Nature.* 2020;581: 465–469. doi:10.1038/s41586-020-2196-x
30. Kohmer N, Rabenau HF, Hoehl S, Kortenbusch M, Ciesek S, Berger A. Comparative analysis of point-of-care, high-throughput and laboratory-developed SARS-CoV-2 nucleic acid amplification tests (NATs). *J Virol Methods.* 2021;291: 114102–114102. doi:10.1016/j.jviromet.2021.114102
31. Owusu D, Pomeroy MA, Lewis NM, Wadhwa A, Yousaf AR, Whitaker B, et al. Persistent SARS-CoV-2 RNA Shedding without Evidence of Infectiousness: A Cohort Study of Individuals with COVID-19. *J Infect Dis.* 2021;224: jiab107-. doi:10.1093/infdis/jiab107
32. Aupaix A, Lazarova E, Chemais M. A brief performance evaluation and literature review of Abbott ID Now COVID-19 rapid molecular-based test. *J Virol Methods.* 2021;298: 114293. doi:10.1016/j.jviromet.2021.114293
33. Thwe PM, Maiyo E, Ren P. Abbott ID now COVID-19 assay performance: a year in review. *Diagn Micr Infec Dis.* 2021;101: 115536–115536. doi:10.1016/j.diagmicrobio.2021.115536
34. Harrington A, Cox B, Snowdon J, Bakst J, Ley E, Grajales P, et al. Comparison of Abbott ID Now and Abbott m2000 Methods for the Detection of SARS-CoV-2 from Nasopharyngeal and Nasal Swabs from Symptomatic Patients. *J Clin Microbiol.* 2020;58: e00798-20. doi:10.1128/jcm.00798-20
35. Mitchell SL, George KSt. Evaluation of the COVID19 ID NOW EUA assay. *J Clin Virol.* 2020;128: 104429–104429. doi:10.1016/j.jcv.2020.104429
36. Basu A, Zinger T, Inglima K, Woo K, Atie O, Yurasits L, et al. Performance of Abbott ID Now COVID-19 Rapid Nucleic Acid Amplification Test Using Nasopharyngeal Swabs Transported in Viral Transport Media and Dry Nasal Swabs in a New York City Academic Institution. *J Clin Microbiol.* 2020;58: e01136-20. doi:10.1128/jcm.01136-20
37. Deslandes V, Clark E, Thiruganasambandamoorthy V, Desjardins M. Implementation of the Abbott ID Now COVID-19 assay at a tertiary care centre: a prospective pragmatic implementation study during the third wave of SARS-CoV-2 in Ontario. *Diagn Micr Infec Dis.* 2021; 115609. doi:10.1016/j.diagmicrobio.2021.115609

38. Kissler SM, Fauver JR, Mack C, Tai CG, Breban MI, Watkins AE, et al. Viral Dynamics of SARS-CoV-2 Variants in Vaccinated and Unvaccinated Persons. *New Engl J Medicine*. 2021;385: NEJMc2102507. doi:10.1056/nejmc2102507
39. Pouwels KB, Pritchard E, Matthews PC, Stoesser N, Eyre DW, Vihta K-D, et al. Effect of Delta variant on viral burden and vaccine effectiveness against new SARS-CoV-2 infections in the UK. *Nat Med*. 2021;27: 2127–2135. doi:10.1038/s41591-021-01548-7
40. Pajon R, Paila YD, Girard B, Dixon G, Kacena K, Baden LR, et al. Initial Analysis of Viral Dynamics and Circulating Viral Variants During the mRNA-1273 Phase 3 COVE Trial. *Medrxiv*. 2021; 2021.09.28.21264252. doi:10.1101/2021.09.28.21264252
41. Ke R, Martinez PP, Smith RL, Gibson LL, Achenbach CJ, McFall S, et al. Longitudinal analysis of SARS-CoV-2 vaccine breakthrough infections reveal limited infectious virus shedding and restricted tissue distribution. *Medrxiv*. 2021; 2021.08.30.21262701. doi:10.1101/2021.08.30.21262701
42. Islam MS, Parvin S, Khan MMR, Hossain GJ, Zahiruddin A, Alam MJ, et al. Asymptomatic RT-PCR positive COVID-19 patients in orthopaedic pre-operative evaluation during the peak of the second wave. *Clin Infect Pract*. 2022; 100131. doi:10.1016/j.clinpr.2021.100131
43. Oran DP, Topol EJ. The Proportion of SARS-CoV-2 Infections That Are Asymptomatic. *Ann Intern Med*. 2021; M20-6976. doi:10.7326/m20-6976
44. Jefferson T, Spencer EA, Maltoni S, Brassey J, Onakpoya IJ, Rosca EC, et al. Viral cultures, PCR Cycle threshold values and viral load estimation for COVID-19 infectious potential assessment in transplant patients: systematic review - Protocol Version 30 December 2021. *Medrxiv*. 2022; 2021.12.30.21268509. doi:10.1101/2021.12.30.21268509
45. Bock MJ, Kuhn MA, Chinnock RE. COVID-19 diagnosis and testing in pediatric heart transplant recipients. *J Hear Lung Transplant*. 2021;40: 897–899. doi:10.1016/j.healun.2021.06.009
46. Spinelli MA, Lynch KL, Yun C, Glidden DV, Peluso MJ, Henrich TJ, et al. SARS-CoV-2 seroprevalence, and IgG concentration and pseudovirus neutralising antibody titres after infection, compared by HIV status: a matched case-control observational study. *Lancet Hiv*. 2021;8: e334–e341. doi:10.1016/s2352-3018(21)00072-2
47. Stemler J, Kramer T, Dimitriou V, Wieland U, Schumacher S, Sprute R, et al. Mobile PCR-based surveillance for SARS-CoV-2 to reduce visiting restrictions in nursing homes during the COVID-19 pandemic: a pilot study. *Infection*. 2021; 1–10. doi:10.1007/s15010-021-01716-4