

Supplementary

Supplementary 1: Spearman's rank coefficient and Bland-Altman Bias mean difference for all sample types.

Sample type	N	Spearman r (95% CI)	P value	Bland-Altman Bias (SD)
Serum	37	0.9915 (0.9829-0.9958)	<0.0001	-0.0248 (0.0708)
EDTA	37	0.9835 (0.9670-0.9918)	<0.0001	-0.0340 (0.0855)
Sodium citrate	37	0.9802 (0.9604-0.9901)	<0.0001	0.0262 (0.0696)
Day 1 fridge	38	0.8810 (0.7813-0.9368)	<0.0001	-0.0018 (0.1224)
Day 1 RT	35	0.9025 (0.8140-0.9500)	<0.0001	0.0574 (0.1965)
Day 3 RT	31	0.9739 (0.9460-0.9875)	<0.0001	-0.0515 (0.1601)
Day 5 RT	36	0.9710 (0.9435-0.9853)	<0.0001	-0.0677 (0.1698)
Day 7 RT	34	0.9646 (0.9297-0.9823)	<0.0001	-0.0755 (0.2014)
DBS	33	0.9485 (0.8974-0.9745)	<0.0001	0.0195 (0.1255)
Total capillary	207	0.9234 (0.9016-0.9405)	<0.0001	-0.00284 (0.1914)

Supplementary 2: Bland-Altman mean difference comparison of OD ELISA results for matched capillary and venous blood samples (collected in lithium heparin treated tubes) (dashed lines represent 95% limits of agreement (-0.3779 – 0.3722))

